



MagPurix[®] Forensic DNA Extraction Kit (ZP02010)

Instructions for Use (Handbook)



Version: 1.9



48



For *in vitro* diagnostic use



ZINEXTS LIFE SCIENCE CORP.
16F., No. 93, Sec. 1, Xintai 5th Rd.,
Xizhi Dist., New Taipei City 221416,
Taiwan (R.O.C.)

Obelis s.a.
Bd Général Wahis 53
1030 Brussels Belgium
Tel: +(32) 2 732-59-54
Fax: +(32) 2 732-60-03
mail@obelis.net



Read and follow these Instructions for Use prior to using this product. The latest revision of this document can be found at www.zinexts.com

Contents

Intended Use	3
Introduction	3
Kit Contents and Storage	3
Materials Required But Not Provided	4
Warnings and Precautions	5
Purification Principle	6
Before Starting	7
Preparation of sample materials	7
Procedure of MagPurix System	10
Purification Protocol - MagPurix® series	10
Purification Protocol - MagPurix® EVO series	11
Troubleshooting	12
Related Products	13
References	13
Symbols	14
Limited Product Warranty	14
Revision History	15

Intended Use

MagPurix® Forensic DNA Extraction Kit provides reagents for fully automated and simultaneous purification of DNA from human whole blood, dried blood, hair root, tissues, saliva and other forensic samples with the MagPurix system.

The product is intended to be used by professional users, such as technicians and physicians who are trained in molecular biology techniques.

Introduction

Product Name	MagPurix® Forensic DNA Extraction Kit
Catalogue Number	ZP02010
Product Overview	The MagPurix® Forensic DNA Extraction Kit is designed to extract of DNA from human whole blood, dried blood, hair root, tissues, saliva and other forensic samples. The kit uses unique magnetic ZiBeads® technology and in combination with MagPurix® series automatic instruments, achieved superior product quality, consistent and high product yield and reproducible results. The final product is suitable for a wide range of diagnostic and research applications, such as sequencing, genotyping, qPCR, ddPCR and NGS assays.
Applicable Instrument Model	All MagPurix® Instrument
Display Protocol Name on The Instrument	2010 FORENSIC DNA
Applicable Instrument Firmware	Check and download the latest firmware from www.zinexts.com
Processing Time	MagPurix® 12 series 40-60 minutes MagPurix® 24 series 40-65 minutes MagPurix® EVO series 40-45 minutes

Kit Contents and Storage


Shipping and Storage	The kit is shipped at room temperature. Upon receipt, store the kit at room temperature. All kit components are stable when stored properly until the expiration date shown on the kit box.	
Kit Content	The components supplied in the kit are listed below. Sufficient reagents are supplied to perform 48 purifications.	
	Contents	Amount
	1 Reagent Cartridge	48 pcs (6x8)
	2 Reaction Chamber	48 pcs (6x8)
	3 Tip Holder	48 pcs (6x8)
	4 Piercing Pin	50 pcs
	5 Filter tip	50 pcs

6 Sample Tube (2 ml)	50 pcs
7 Elution Tube (1.5 ml)	50 pcs
Filter Column	50 pcs
Collection Tube	50 pcs
Proteinase K, 10 mg/mL (1 ml)	1 pc
BL2 Buffer (25 ml)	1 pc
Barcode sticker (EVO only)	50 pcs

Reagent
Cartridge
Contents

Each Reagent Cartridge has 10 positions with 10 sealed well.
Positions 1-10 contain wells filled reagents for this protocol

Reagent	Well No.
Empty	1
Lysis Buffer 3	2
Binding Buffer 1	3
Magnetic Bead Solution	4
Washing Buffer 1B	5
Washing Buffer A	6
Washing Buffer B	7
Elution Buffer 1	8
Elution Buffer 2	9
Empty	10



Materials Required But Not Provided

The following general laboratory equipment and consumables are required to perform the extraction. All laboratory equipment should be installed, calibrated, operated, and maintained according to the manufacturer's recommendations. The following tables display required and special equipment along with the list of consumables.

Item
MagPurix® series instrument
1.5 or 2.0 ml microcentrifuge tubes
Pipettes and filter tips
Phosphate-buffered saline (PBS, may be required for diluting samples)
Optional: Plastic consumables, DNase-free RNase A (to minimize RNA content)

Warnings and Precautions

For *in vitro* diagnostic use only. Read all the instructions carefully before using the kit. Use of this product should be limited to trained personnel in the techniques of DNA purification. Strict compliance with the user manual is required for optimal results. Attention should be paid to expiration dates printed on the box and labels of all components. Do not use a kit after its expiration date.

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in convenient and compact PDF format at


<http://www.zinexts.com/download.php?lang=en&tb=1&cid=7> where you can find, view, and print the SDS for each kit and kit component.

Please report any serious incident occurred in relation to the device to your local representative/agent or the manufacturer, and to the competent authority of your country/state.



CAUTION: DO NOT add bleach or acidic solutions directly to the sample preparation waste.

Purification Principle



A Perform certain pretreatment process before extraction.

B Transfer sample to extraction directly.

Before Starting

Preparation of sample materials

The purification procedure is optimized for the use of appropriate volume of human whole blood, dried blood, hair root, tissues, saliva and other forensic samples.

Whole Blood (fresh or frozen)	NOTE: To extract DNA from whole blood samples, please select and refer to MagPurix® Blood DNA Extraction Kit 200 (ZP02001).
Clotted/Dried Blood	<ol style="list-style-type: none">Cut the blood-contained range out, transfer pieces to a 1.5 ml microcentrifuge tube.Add 400 µl BL2 Buffer and 20 µl Proteinase K to the sample.Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer.Transfer all the sample into a Filter Column sitting in Collection Tube.Short spin at 500 x g, 1 minute to collect the clear flow-through in Collection Tube.Transfer 400 µl into each Sample Tube.
Forensic Surface and Contact Swabs	<ol style="list-style-type: none">Allow the swab or brush to air-dry for at least 2 hours after collection.Carefully cut or break off the end part of the swab or brush into a 1.5 ml microcentrifuge tube using an appropriate tool (e.g., scissors).Add 200-400 µl BL2 Buffer to the sample.Add 20 µl Proteinase K, vortex for at least 10 seconds.*If processing brush samples, centrifuge the tube briefly (at 10,000 x g for 30 seconds) to force the brush down to the bottom of the tube.Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer.Pre-filter the digested lysate using a Filter Column to remove residual debris.Short spin at 500 x g, 1 minute to collect the clear flow-through in Collection Tube.Transfer 200-400 µl into each Sample Tube.
Hair Root	<p>Method 1</p> <ol style="list-style-type: none">Place the hair sample in a 1.5 ml microcentrifuge tube.Add 200 µl BL2 Buffer to the sample.Add 20 µl Proteinase K and 10 µl 1M DTT solution*, mix thoroughly by vortexing for at least 10 seconds.Incubate at 56°C for 15 minutes to 6 hours, vortex several times during incubation or place the sample in a thermomixer.(optional) Add extra 10 µl Proteinase K and 10 µl DTT and incubate at 56°C until the hair samples are completely dissolved.Spin the tube to collect drops from inside the lid.Pre-filter the digested lysate using a Filter Column to remove residual debris.Short spin at 500 x g, 1 minute, to collect the clear flow-through in Collection Tube.Transfer 200 µl into each Sample Tube. <p>*Prepare 1 M DTT solution before processing the protocol (1 M is about 15% DTT (m/v)).</p>

Method 2

- a. Place the hair sample in a 1.5 ml micro centrifuge tube.
- b. Add 200 μ l BL2 Buffer to the sample.
- c. Add 20 μ l Proteinase K, mix thoroughly by vortexing for at least 10 seconds.
- d. Incubate at 56°C for at least 15 minutes or overnight, vortex several times during incubation or place the sample in a thermomixer.
- e. Spin the tube to collect drops from inside the lid.
- f. Pre-filter the digested lysate using a Filter Column to remove residual debris.
- g. Short spin at 500 x g, 1 minute to collect the clear flow-through in Collection Tube.
- h. Transfer 200 μ l into each Sample Tube.

Human Tissues

- a. Place tissue sample into a 1.5 ml micro centrifuge tube.
- b. Add 200-400 μ l BL2 Buffer and 20 μ l Proteinase K to the sample, mix thoroughly by vortexing for 10 seconds.
- c. Incubate at 56°C for at least 2 hours, *vortex several times during incubation or place the sample in a thermomixer.
- d. *Incubation for longer time (e.g., overnight) will not interfere the extraction.
- e. Spin the tube to collect drops from inside the lid.
- f. Pre-filter the digested lysate using a Filter Column to remove residual debris.
- g. Short spin at 500 x g, 1 minute to collect the clear flow-through in Collection Tube.
- h. Transfer 200-400 μ l into each Sample Tube.

Saliva

- a. Place up to 50 μ l saliva in a 1.5 ml microcentrifuge tube.
- b. Add 200 μ l BL2 Buffer to the sample.
- c. Add 20 μ l Proteinase K, and mix thoroughly by vortexing for 10 seconds.
- d. Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer.
- e. Spin the tube to collect drops from inside the lid.
- f. Transfer 200 μ l into each Sample Tube.

Sperm Stains

- a. Place the forensic sample in a 1.5 ml microcentrifuge tube.
- b. Add 200-400 μ l BL2 Buffer to the sample.
- c. Add 20 μ l Proteinase K and mix thoroughly by vortexing for 10 seconds.
- d. Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer.
- e. Spin the tube briefly to collect drops from inside the lid.
- f. Pre-filter the digested lysate using a Filter Column to remove residual debris and mucus.
- g. Short spin at 500 x g, 1 minute to collect the clear flow-through in Collection Tube.
- h. Transfer 200-400 μ l into each Sample Tube.

Chewing Gum

- a. Place the chewing gum sample in a 1.5 ml microcentrifuge tube.
 - b. Add 200 μ l BL2 buffer to the sample.
 - c. Add 20 μ l Proteinase K, and mix thoroughly by vortexing for 10 seconds.
-

	<ul style="list-style-type: none"> d. Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer. e. Spin the tube briefly to collect drops from inside the lid. f. Transfer 200 µl into each Sample Tube.
Cigarette Butts	<ul style="list-style-type: none"> a. Place the cigarette butt sample in a 1.5 ml micro centrifuge tube. b. Add 200 or 400 µl BL2 Buffer to the sample. c. (Check if the sample has absorbed BL2 Buffer, add more BL2 Buffer to the sample if necessary). d. Add 20 µl Proteinase K, and mix thoroughly by vortexing for 10 seconds. e. Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer. f. Spin the tube briefly to collect drops from inside the lid. g. Transfer 200 µl into each Sample Tube.
Stamps, Envelopes	<ul style="list-style-type: none"> a. Add 200-400 µl BL2 Buffer to the sample. (Check if the sample has absorbed BL2 Buffer, add more BL2 Buffer to the sample if necessary). b. Add 20 µl Proteinase K, and mix thoroughly by vortexing for 10 seconds. c. Incubate at 56°C for 15 minutes, vortex several times during incubation or place the sample in a thermomixer. d. Spin the tube briefly to collect drops from inside the lid. e. Transfer 200-400 µl to each Sample Tube.

Note:

DNA quality will decrease with time or after multiple freeze–thaw cycles. For longer storage time, samples should be frozen at -20°C or lower and avoid freeze–thaw cycles. Thaw samples at room temperature (15-25°C) and process the sample immediately after equilibration to room temperature. **Do not** refreeze sample after thawing. If precipitation is visible in sample, centrifuge at 6,800 x g for 3 minutes and transfer supernatant to a new tube without disturbing the precipitate, and immediately start the purification procedure.

The suggested starting material and elution volume range for each nucleic acid extraction.

Sample type	Starting material per sample	Elution Volume
Whole Blood (fresh or frozen)	100-400 µl	50-300 µl (EVO 50-200 µl)
Clotted/Dried Blood	400 µl or 3 card punches* * Using a single-hole paper punch to cut 3 mm (1/8 inch) diameter of punches from a dried blood spot.	
Forensic Surface and Contact Swabs	200-400 µl	
Hair Root	100-400 µl/two or three 0.5-1 cm from the root ends of plucked hair samples	
Human Tissues	100-400 µl/up to 40 mg tissue	
Saliva	100-400 µl/50 µl or more volume of saliva	
Sperm Stains	100-400 µl/5-10 µl or 1 cm ² of the forensic sample	
Chewing Gum	200-400 µl/up to 40 mg of chewing gum cut into small pieces	

Cigarette Butts	200-400 µl/approximately 1 cm ² paper from the end of the cigarette or filter	
Stamps, Envelopes	200-400 µl/a 0.5-2.5 cm ² piece of postage stamp or envelope	

Procedure of MagPurix System

Workflow of MagPurix operation

Place the cartridge and plastic consumables on the MagPurix instrument



Select the protocol and setup the condition



Follow onscreen message for worktable setup



Start the protocol



Collect elution product *



UV decontamination

* Output the bench record (option)

Note: Perform all steps at room temperature (20-25°C) unless otherwise notified.

Purification Protocol - MagPurix[®] series


1	Turn on the Instrument	a. Turn on the power switch and wait for the screen to turn on.
2	Load new Consumable(s) and Cartridge(s)	a. Open the door and remove the Sample Rack from the instrument. b. Load 1 Reagent Cartridge, and all plastic disposables (2 Reaction Chamber, 3 Tip Holder, 4 Piercing Pins, 5 Filtered Tips and other components presented in the kit intended to use). Place 6 Sample Tubes and 7 Elution Tubes into the Sample Rack.
3	Load the Samples	a. Transfer appropriate volume of sample into each Sample Tube on the Sample Rack. b. Put the Sample Rack back into the instrument and close the door.
4	Program Set up	a. Scan the protocol barcodes to select the purification protocol, sample volume and elution volume.
5	Start Extraction	a. Follow the instructions displayed on the screen to double-check the operating steps being completed before program running. b. Press “ENTER” to start the experiment. Instrument will run the protocol program automatically until the whole process is completed. c. At the end of the run (approximately 12 series 40-60 minutes , 24 series 40-65 minutes), instrument alarms briefly.

-
- | | | |
|----------|---------------------------|--|
| 6 | Collect the Elution tubes | <ul style="list-style-type: none"> a. Open the instrument door. b. Collect the elution tubes containing the purified nucleic acids. c. The purified nucleic acids are ready for immediate use. Store the purified nucleic acids at 4°C (short-term, less than 10 days) or aliquot and store at -70°C (long-term) before performing downstream analysis. d. Discard the used cartridges and all plastic consumables into biohazard waste. *Do not reuse the cartridges. e. If you are not using the instrument immediately, place the Sample Rack back to the workplace, close the instrument door and press “Start” button for 2 seconds to enter sleep mode. Moreover, if the instrument will not be used in a long time, please turn off the power switch. |
|----------|---------------------------|--|
-

Purification Protocol - MagPurix® EVO series

-
- | | | |
|----------|---|--|
| 1 | Turn on the Instrument | <ul style="list-style-type: none"> a. Turn on the power switch and wait for the screen to turn on. a. Login the instrument and enter the Home Page. |
| 2 | Load new Consumable(s) and Cartridge(s) | <ul style="list-style-type: none"> a. Open the door and remove the Sample Rack from the instrument. b. Open the Tip-Holder Lid. c. Load 1 Reagent Cartridge and all plastic disposables (2 Reaction Chamber, 3 Tip Holder, 4 Piercing Pins, 5 Filtered Tips and other components presented in the kit intended to use). d. Close the Tip-Holder Lid. e. Paste the barcode stickers on Elution Tubes. a. Place 6 Sample Tubes and 7 Elution Tubes into the Sample Rack. |
| 3 | Load the Samples | <ul style="list-style-type: none"> a. Transfer appropriate volume of sample into each Sample Tube on the Sample Rack. a. Put the Sample Rack back into the instrument and close the door. |
| 4 | Program Set up | <ul style="list-style-type: none"> a. Select the appropriate protocol program on the instrument. Press NEXT. b. Select the appropriate Sample Volume and Elution Volume and press NEXT. c. Press the number button to select the right Sample Numbers. d. Scan/Edit each primary Sample ID directly. After finished, press NEXT. e. Scan/Edit each Elution Tube ID directly. After finished, press NEXT. f. Scan Reagent Cartridge Barcode. Press NEXT. *If the cartridge is expired, the next step cannot be performed. a. Follow the instructions on the screen to double-check the operating steps being completed before running the program. Press NEXT. |
| 5 | Start Extraction | <ul style="list-style-type: none"> a. Check “PROGRAM CONFIRMATION” on the screen. b. Press “START” to start the experiment. Instrument will run the protocol program automatically until the whole process is completed. c. At the end of the run (approximately 40-45 minutes), instrument alarms briefly and the screen indicates “PROGRAM FINISH”. d. If you want to perform the same protocol, press “RERUN” to perform the same experiment. If you do not need to re-run the experiment, |
-

press the function button “ HOME” to exist the experiment mode.

- 6** Collect the Elution tubes
- Open the instrument door.
 - Collect the elution tubes containing the purified nucleic acids.
 - The purified nucleic acids are ready for immediate use. Store the purified nucleic acids at 4°C (short-term, less than 10 days) or aliquot and store at -70°C (long-term) before performing downstream analysis.
 - Discard the used cartridges and all plastic consumables into biohazard waste. ***Do not reuse the cartridges.**
 - If you are not using the instrument immediately, please put the Sample Rack back into the instrument, close the instrument door, and press the “ POWER” function button to enter sleep mode. If the instrument will not be used in a long time, please turn off the power switch.

Troubleshooting

***This table is helpful for solving common problem. If you need other technical support, please contact Zinexts team (<http://www.zinexts.com/index.php?lang=en>) or your distributor.**

Problem	Possible Cause	Comments and suggestions
Poor DNA quality or yield	Deterioration or contamination of reagents.	Please ensure that the kit reagents are still in the effective using period before use. Discard any kit reagent that shows discoloration or evidence of microbial contamination.
	Kit stored under non-optimal conditions	Store kit at 15-25°C at all time after arrival. If either reagent or buffer precipitate upon shipping in cold weather or during long-term storage, dissolve precipitates by gently warming and stirring the solution. Please do not freeze the Reagent Cartridges.
	Insufficient sample input	DNA yield depends on the sample type and the number of nucleated cells in the sample. Please proportionally adjust the total input amount of sample to increase the DNA yield.
	Too much of elution buffer was used	The elution volume can be reduced proportionally.
	The eluate of final product(s) is not enough.	Please collect issue information and provide it to your Support Representative/Technical Support as soon as possible.

Clogging issue	Too much sample material was used.	Decrease the input amount of sample material or dilute your sample.
No results in downstream analysis	No signal/The PCR was inhibited.	Using appropriate controls for analysis. Check the positive control, negative control, water (NTC) and internal control to clarify the possible causes.
Instrument malfunction/abnormal sound	Abnormal consumables: 1. Deformed Filtered Tips 2. Deformed Reaction Chamber 3. Deformed Tip Holder	Please replace the batch with normal consumables.
	Abnormal action of instrument: 1. Inaccurate correction value 2. Spare parts or components damaged	Please collect issue information (videos and pictures) and provide it to your Support Representative/Technical Support as soon as possible to calibrate or replace any other damaged or worn parts.

Related Products












Product Name	Cat. no.
MagPurix® Blood DNA Extraction Kit 200	ZP02001
MagPurix® Blood DNA Extraction Kit 1200	ZP02002
MagPurix® Viral NA Extraction Kit	ZP02003
MagPurix® Tissue DNA Extraction Kit	ZP02004
MagPurix® Cultured Cell DNA Extraction Kit	ZP02005
MagPurix® Bacterial DNA Extraction Kit	ZP02006
MagPurix® HPV DNA Extraction Kit	ZP02007
MagPurix® TB DNA Extraction Kit	ZP02008
MagPurix® FFPE DNA Extraction Kit	ZP02009
MagPurix® Forensic DNA Extraction Kit	ZP02010
MagPurix® Viral Pathogen DNA Extraction Kit A	ZP02011
MagPurix® Viral Pathogen DNA Extraction Kit B	ZP02012
MagPurix® Viral RNA Extraction Kit	ZP02013
MagPurix® Plant DNA Extraction Kit	ZP02014
MagPurix® Total RNA Extraction Kit	ZP02015
MagPurix® Viral NA Extraction Kit LV	ZP02016
MagPurix® CFC DNA Extraction Kit	ZP02017
MagPurix® cfDNA Extraction Kit Plus	ZP02024
MagPurix® cfDNA Extraction Kit LV	ZP02025
MagPurix® Coronavirus RNA Extraction Kit	ZP02027

References

- Tan SC *et al.* J Biomed Biotechnol. (2009)

Symbols

The following symbols are used on labels and in Instructions for Use (IFU), in compliance with EN ISO 15223-1 standard.

Symbol	Explanation
	CE mark
	For In Vitro Diagnostic Use
	Catalogue number
	Lot/Batch number
	Sufficient for [n] samples
	Instructions for Use
	Expiry date
	Storage temperature (15°C - 25°C)
	Manufacturer
	European Authorized Representative
	Caution

Limited Product Warranty

Zinexts Life Science is committed to provide customers with high-quality products and services. Our goal is to ensure that every customer is 100% satisfied with our products and services. If you have any question or concerns, contact our Technical Support Representatives.

Zinexts Life Science guarantees the performance of all products according to the specifications stated in our product literature. The purchasers/users must determine the suitability of the product for their particular use. We reserve the right to change, alter, or modify any product to enhance its performance and design.

This warranty limits Zinexts Life Science Corporation's liability only to the cost of the product. No warranty is granted for products beyond their listed expiration date. No warranty is applicable unless all product components are stored and used in accordance with instructions.

Revision History

Version	Date	Description
1.9	11. Aug. 2022	List of IVD symbols added